


Betaine

Method:

Column	Ultisil XB-NH2, 5 μ m, 4.6 \times 250mm (P/N 00204-31043)
Mobile Phase	Acetonitrile:Water=75:25
Detection	192nm
Temperature	30°C
Flow Rate	1.0ml/min
Injection Volume	10 μ l
Note	/

Chromatogram and Data:


Reference Solution



#	Rt	Area	Height	Plates(USP)	Tailing Factor	Resolution
1	12.130	14517537	615315	5752	0.954	--

Sample Solution

Unless otherwise stated, the results shown in this test report refer only to this sample tested. The report can not be copied without the permission of Welch Materials, Inc.



#	Rt	Area	Height	Plates(USP)	Tailing Factor	Resolution
1	12.106	522495	31588	12526	1.022	--

Unless otherwise stated, the results shown in this test report refer only to this sample tested. The report can not be copied without the permission of Welch Materials, Inc.