


Sulfogaiacol Pentoxyverine

Method:

Column	Ultisil® AQ-C18 4.6×250 mm, 5 µm	P/N: 00207-31043
Mobile Phase	0.05mM Potassium heptane sulfonate and 0.5ml/L triethylamine, adjust pH into 3.0 with phosphoric acid: Methanol=50:50	
Detection	218 nm	
Temperature	30 °C	
Flow Rate	1.0 ml/min	
Injection Volume	5 µl	

Chromatogram and Data:

1. Sample



ID.	Rt.	Height.	Area.	half peak width.	Resolution.	Concentration.	Th. Plate.	Tailing.	Components.
1	3.198	862	4726.6	1.784	5.15	0.00	7675	0.95	
2	3.407	851	4510.7	1.703	4.98	1.45	9318	1.22	
3	3.707	32686	232522.9	87.783	6.69	1.81	6124	1.53	
4	14.623	363	15667.0	5.915	40.57	16.31	2589	1.37	
5	18.807	128	7455.6	2.815	54.75	3.10	2352	0.97	
		34890	264682.8	100.0000					Σ :

Unless otherwise stated, the results shown in this test report refer only to this sample tested. The report can not be copied without the permission of Welch Materials, Inc.